

## ***Clostridium difficile* Associated Diarrhea**

### **To the Editor:**

*Clostridium difficile*-associated diarrhea and colitis is frequently noted in patients receiving antibiotic therapy in hospitals<sup>1</sup> and chronic-care facilities.<sup>2</sup> The organism is usually acquired nosocomially through human contact or from the hospital environment that is frequently contaminated with *C difficile* spores.<sup>3</sup> Immunoblot typing has been used to demonstrate nosocomial infection.<sup>4</sup>

As part of an investigation of an outbreak of *C difficile*-associated diarrhea, we studied the possible role of rectal thermometers as a source of nosocomial transmissions. Over a one-month interval, a total of 72 electronic rectal thermometers and 16 electronic oral thermometers were tested for *C difficile*. The portion of the handle nearest the base of the disposable plastic temperature probe sheath was sampled using a sterile moist swab that was then inoculated into chopped meat broth containing cefoxitin (40 µg/ml) to suppress contaminants, and sodium taurocholate (1 mg/ml) to promote spore germination. After 24 hours

of incubation, the broths were screened for *C difficile* antigen (Marion Latex Test), and those that tested positive were further evaluated by *C difficile* isolation and identification.

Fifteen of 72 rectal thermometers (20.8%) were positive for *C difficile* antigen, and of these, seven were confirmed by isolation and identification of the organism. The remaining eight specimens were overgrown with enterococcus, which interfered with the attempted isolation of *C difficile*. None of the oral thermometers were positive for *C difficile* by the antigen screen.

Electronic rectal thermometers, which are used extensively in many institutions, are designed for multiple patient use. Our study shows that the handles often become contaminated with *C difficile* spores. Although the temperature probe is covered by a disposable plastic sheath that is changed between each patient, the newly installed sheath may be easily contaminated by hand carriage of spores from the handle of the thermometer.

Insertion of a temperature probe carrying *C difficile* spores directly into the rectum might be a mechanism of nosocomial transmission of the organisms between

patients. Thorough cleaning of the handle of the thermometer between patient use may help reduce the transmission of *C difficile* by this route.

**S. Brooks, PhD;  
R. Real, AAS;  
M. Kramer, MD,  
M. Adachi, MD, ScD**

Kingsbrook Jewish Medical Center  
Brooklyn, New York

### **REFERENCES**

1. Bartlett JG, Chang TW, Gurwith M, Gorbach SL, Onderdonk AB. Antibiotic-associated pseudomembranous colitis due to toxin producing *Clostridia*. *N Engl J Med*. 1978;298:531-534.
2. Bender BS, Bennett R, Laughon BE, et al. Is *Clostridium difficile* endemic in chronic care facilities? *Lancet*. 1986;2:11-13.
3. Savage AM, Alford RH. Nosocomial spread of *Clostridium difficile*. *Infect Control*. 1983;4:31-33.
4. McFarland LV, Mulligan ME, Kwok RYY, Stamm WE. Nosocomial acquisition of *C difficile* infections. *N Engl J Med*. 1989;320:204-210.

**Letters to the Editor should be addressed to INFECTION CONTROL AND HOSPITAL EPIDEMIOLOGY Editorial Offices, C41 General Hospital, University of Iowa Hospitals and Clinics, Iowa City, IA 52242. All letters must be typed, double spaced, and may not exceed four pages nor include more than one figure or table. The editors reserve the right to edit for purposes of clarity or brevity.**