

We appreciate the response to this publication feature and welcome all contributions. Contributions may be sent to our Technical Editor Phil Oshel, oshel1pe@cmich.edu

## A Simple Cleaning Method for Penning Gauges

Valery Ray

PBS&T, MEO Engineering Co., Methuen, MA  
www.partbeamsystech.com

The usual failure mode of cold cathode ionization gauges (CCIG) in SEM and FIB tools is dielectric deposition from hydrocarbon oils and precursor gases on electrodes of the sensor tube. Removal of these deposits fully restores the functionality of the gage. So far, I have cleaned hundreds (literally) of Edwards and Varian CCIG sensors by following this procedure:

- 1) Dismount the CCIG from the tool, remove the magnet, and carefully take apart the sensor tube.
- 2) Soak all the metal parts of the sensor tube overnight in a 50:50 solution of Micro 90™ and deionized or distilled water (Micro 90 is available from various science-supply houses). If deionized water is not available, then supermarket distilled water works. Avoid tap water.
- 3) In the morning, sonicate all metal parts for 10 minutes in the same solution in which they were soaking. At this point all deposits will be removed from the electrodes, resulting in nice, shiny, metal surfaces.
- 4) Thoroughly rinse the parts, dry them (use a hot plate or oven to speed up the drying), and put the sensor tube together.

Micro 90™ cleaning is very gentle, does not scratch electrodes, and takes a minimum of personnel time - other things can be done while the gauge is soaking or drying. Cleaning by hand-polishing of the electrodes will also work, but this takes personnel time.

A good idea would be to pre-clean one or two old gauges, which may be lying around the lab, and replace the gauge on the instrument as needed. Store the pre-cleaned gauges loosely sealed in aluminum foil in a desiccator.

Soaking in Micro 90™ is generally helpful for cleaning contamination from metal surfaces. I use this same procedure for cleaning extractor and suppressor electrodes of FIB columns, and also have tried it on the Wehnelt of a tungsten-filament SEM. With column parts, some very light polishing by 1 μm paste is usually needed in the end to remove the most stubborn deposits.

If parts are not heavily contaminated and work done carefully (gloves and tweezers), no degreasing is needed after Micro 90™, as all of the grease is removed. Wear gloves to protect hands! Micro 90™ is much stronger than regular liquid detergent and will "degrease" skin.

Caution - I never tried Micro 90™ soaking of parts with plated surfaces, chances are that plating might go away as easily as the contamination. ■

### Editor's Note ...

What do you think about MT publishing detailed protocol articles, like the one by Hazelton earlier in this issue and these M-101 notes? Please send me an email with your thoughts. ...

microscopytoday@tampabay.rr.com

## Negative Stains/Staining 2.5 mM Phosphotungstic Acid, 25μg/ml Bacitracin, pH 7.0

Paul R. Hazelton

University of Manitoba, Winnipeg, Manitoba, Canada  
paul\_hazelton@umanitoba.ca

1. Combine:
  - a. 406 mg Phosphotungstic Acid  $\{(W_{12}O_{41})_2 \cdot 2H_3PO_4 \cdot 48H_2O\}$ , MW 6498.93; and
  - b. Glass redistilled water to volume 20.0 mL.
- 1 Stir until dissolved.
- 2 Adjust pH to 7.0 with NaOH.
- 3 Add 0.625 mg Bacitracin.
- 4 Stir until dissolved.
- 5 Bring to final volume 25.0 ml with glass redistilled water.
- 6 Filter through a 0.2 μm filter and store at 4°C.

NOTE: PTA goes into solution very readily but Bacitracin will not go into solution at low pH. The initial pH of a PTA preparation is around 2.0. Therefore, it must be raised to near 7.0 before adding Bacitracin. The operative staining molecule is the  $W_{12}O_{41}$  polyanion, which may pass between that structure and  $WO_4$ . The polyanions have a reported diameter of 0.8-0.9 nm (Hayat and Miller, 1990). The anhydrous density of the stain is reported to be 4.0-4.2. The stain is very stable at 4°C, showing no signs of deterioration in the form of stain precipitate, drop in pH, or deterioration in the image produced in the electron microscope. The purpose of Bacitracin is as a surfactant, enhancing spreading. The Bacitracin molecule is smaller than that of albumin, and is, therefore, preferable as a spreading agent (Gregory and Pirie, 1973). Most samples have enough protein present, other than those in gradients, to negate the need of spreading agents. ■

- 1 Gregory DW and Pirie BJ. 1973. Wetting agents for biological electron microscopy. I. General considerations and negative staining. *J Microsc.* 99:251-5.
- 2 Hayat and Miller, 1990, Negative Staining

## Agar Diffusion

Paul R. Hazelton

University of Manitoba, Winnipeg, Manitoba, Canada  
paul\_hazelton@umanitoba.ca

This procedure is a tremendous improvement upon Kellenberger's original collodion diffusion procedure, where collodion films were cast on agar and the suspension placed on the drop of film, allowed to diffuse through, and the film then floated off on water<sup>3</sup>. The modification was recommended by Anderson and Doane as a method for both concentrating samples onto a grid and dialysing (for lack of a better word) excess salts out. To remove excess salts allow the sample's solute to diffuse into the agar, place sterile filtered water over the sample, and allow the excess salts to diffuse into the agar. They reported that they could see an average of 5 particles/square, and that sensitivity was increased by up to 100×<sup>1</sup>.

1. Preparation of agar plates:
  - a. Prepare 1% agar (Note a) in PBS (Note b).
  - b. Fill wells of a flexible microtitre plate (Note c) approximately 3/4 full with the 1% agar solution.