

USED EQUIPMENT FOR SALE

Cambridge Stereoscan 100 SEM with Backscattered Electron Detector.

In excellent condition, newer turbo-molecular pump, limited usage, under service contract. Call John at (802)656-4504.

NEWPORT CONFOCAL ATTACHMENT, VX-100 Spinning Disk Confocal

adapter, complete for upright microscope. Dealer's demo unit, like-new condition. When new, was \$25,000, asking \$15,000. Call Hitech Instruments, Inc.: (800)-4-HITECH(444-8324) or fax: ((610)353-3317.

MILITARY RESEARCH LAB IS CLOSING.

Military contractor is selling at drastically reduced prices its Sorvall ultramicrotome, refrigerated and benchtop microtomes, sliding microtome, Tissue Tech embedding center, stereo microscopes, Joyce Loebel microdensitometer and LECO sulfur analyzer. For specification sheets, call: (202)544-0836.

Downloading ImageTool

UTHSCSA ImageTool (IT) is a free image processing and analysis program for Microsoft Windows 95 or Windows NT. IT can acquire, display, edit, analyze, process, compress, save and print gray scale and color images. IT can read and write over 22 common file formats including BMP, PCX, TIF, GIF and JPEG. Image analysis functions include dimensional (distance, angle, perimeter, area) and gray scale measurements (point, line and area histogram with statistics). ImageTool supports standard image processing functions such as contrast manipulation, sharpening, smoothing, edge detection, median filtering and spatial convolutions with user-defined convolution masks. IT also has built-in macro capabilities that allow the user to record repetitive tasks and playback saved macros to automate image analysis.

ImageTool was designed with an open architecture that provides extensibility via a variety of plug-ins. Support for image acquisition using either Adobe Photoshop plug-ins or Twain scanners is built-in. Custom analysis and processing plug-ins can be developed using the software development kit (SDK) provided (with source code). This approach makes it possible to solve almost any data acquisition or analysis problem with IT.

ImageTool provides for geometric transformations such as rotate, flip vertical, flip horizontal and magnification up to four levels. All analysis and processing functions are available at any magnification factor. The program is a multiple document interface (MDI) application supporting any number of windows (images) simultaneously.

Spatial calibration is available to indicate real world dimensional measurements such as millimeters, microns, feet, miles, etc. for linear and area. Density or gray scale calibration can be done relative to radiation or optical density (OD) standards.

IT version 1.1 now provides for object analysis and classification with over 20 morphological descriptors such as: area/perimeter, roundness, ferret diameter, compactness, major/minor axis length, centroid and many others. Any of these factors can be used automatically categorized and count objects within the image.

ImageTool ver. 1.1 supports the Data Translation DT3155 frame grabber for Windows NT. Other frame grabber boards will be added in the coming months.

UTHSCSA ImageTool is available via anonymous ftp at ftp:

//maxrad6.uthscsa.edu:

A Few Remarkable TEM Facts

Phil Fraundorf, University of Missouri

What follows is a list of some physical perspectives on the electrons used routinely for transmission electron microscopy. Without knowing it, you may on a daily basis be putting to practical use things, like the wave nature of electrons, that were inconceivable in the early part of this century. In fact, some of the properties of these electrons may be only marginally conceivable today!

Fast Electrons: A back of the envelope calculation for 300 keV electrons gives $\gamma = (300=511)/511 = 1.587$, so that they travel at $w = c^2[1-(1-\gamma^2)^{-1/2}] = 0.777 c$ or (lightyears per inertial year) of elapsed time. However, if we consider traveler (i.e., electron or proper) time for such a speeding electron, this would give that the travel $u = \gamma w = 1.232$ lightyears per traveler year of elapsed time! With this spatial 4-vector velocity well over c , we're dealing with relativity in action! I wonder how many g's of acceleration they experience in the electron gun in order to get up to speed? For more on this subject, you might want to check our browser-interactive relativistic Accel-One problem solver, and the theory pages attached, at <http://newton.umsl.edu/run/index.html>

Lonely Electrons: I think that it was John Armstrong at Caltech who once pointed out to me that the number of microscope beam electrons in your TEM specimen at any one time is so small that the odds of such electrons interfering with each other to form diffraction patterns is quite small. The vertical separation between electrons in the column is w/l , where l is the specimen current, w is the electron inertial velocity, and e is the charge per electron. For an nanoamp of 300 kV electrons, this is $(0.777 \times 3 \times 10^8 \text{ m/s}) \times (1.6 \times 10^{-19} \text{ C/e}) / (10^{-9} \text{ C/s}) = 0.037 \text{ m/e}$. Under some illumination conditions there may be no more than 1 beam electron in the column at a time! Hence diffraction patterns in the TEM are basically formed by individual electrons interfering with themselves! As you know, such interference will occur only if we don't take steps to determine the path of individual electrons through the specimen! If we look too closely at these paths, the diffraction patterns would disappear (cf. Englert et al., *Scientific American*, Dec. 1994, 86-92 on quantum erasure).

Fat Electrons: The transverse coherence widths of electrons which make possible electron phase contrast (HREM) lattice imaging and probably electron holography might also be seen as lateral broadening of individual electron wave-packets via the uncertainty principle, which results because we know too much about their transverse momentum! My intuition tells me that we're talking about lateral wave-function spreads of, say, 15 Angstroms in a LaB₆ HREM to more than 100 Angstroms in field emission gun systems. Are these numbers reasonable? By increasing the spread of electron angles in the incident beam, this transverse coherence width can presumably be decreased (e.g., you want it small for Z-contrast imaging (I think), or varied as in the variable coherence-width strategies of Murray Gibson at U. of I.

Long Electrons: The tight tolerances on high voltage stability and the emitted spread in electron energies means that our uncertainty in the longitudinal momentum of TEM electrons is quite small, and hence again by the uncertainty principle that the wave-packet spread in the direction of motion for TEM electron can be quite large. Distances of, say, 1000 Angstroms come to mind! The associated tight distribution of incident electron energies decreases chromatic and instability damping of fine details in CTEM and HREM images, so that for most applications you may want your electrons "as long as possible". An exception might be in variable-coherence strategies (mentioned above), where shorter electrons might provide sensitivity to shorter-range vertical correlations.

The foregoing thoughts on fast, lonely, fat, and long electrons are not really things I've had time to think much about, but they are interesting, and hence I would enjoy other perspectives on them, as well as suggestions for other "remarkable TEM facts" to add to the list! A "live" draft of this list will be accessible through our scanned Tip & Electron Image Lab page at: <http://newton.umsl.edu/stei-lab/>

★
★ **A CORRECTION:** ★
★ In "Resolutions Considerations for Photomicrography and Photomicroscopy", page 10 in our May 1996 issue, the equation in 4) should read: ★
★ $(2 \text{ Print Resolution})^2 = (2 \text{ Max Print Resolution})^2 + C^2$. The area of the blurr ★
★ circle from the combined diffraction and geometric blurr's is equal to the ★
★ sum of the blurr circle areas. Table 1 is based upon this correct ★
★ relationship. ★
★
